

EARL PET/CT

Accreditation 2021

THIS IS TO CERTIFY THAT THE

University Hospital Olomouc
Department of Nuclear Medicine
Scanner: Siemens Biograph 40 mCT

is granted accreditation according to the requirements and specifications indicated in the EARL PET/CT accreditation manual for the following oncology isotope imaging and accreditation program:

^{18}F standards 1, ^{18}F standards 2

PET/CT scanner accreditation ensures similar performance of PET/CT systems within a multicentre setting by harmonising reconstructions for quantification of PET/CT scan. Accredited PET/CT centers of excellence can compare, exchange and combine PET/CT findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.



Jolanta Kunikowska
President 2021-2022



Ivalina Hristova
Programme Director

